This manuscript is a preprint and has been submitted for publication in **Journal of Glaciology**. It has been revised after peer review but **not yet accepted for publication**. Subsequent versions of this manuscript may differ due to the editorial process. If accepted for publication, the final version will be available via the "Peer-reviewed publication DOI" link on EarthArXiv. We hope you find this paper interesting and would welcome your feedback on it. Please contact Mike Prior-Jones or Liz Bagshaw with your feedback using the email addresses on the following page.

1 Cryoegg: development and field trials of a wireless subglacial probe for deep, fast-moving ice

- 2 Michael Prior-Jones¹, Elizabeth A. Bagshaw¹, Jonathan Lees², Lindsay Clare³, Stephen Burrow³, Mauro
- 3 A Werder⁴, Nanna B. Karlsson⁵, Dorthe Dahl-Jensen^{6,7}, Thomas R. Chudley⁸, Poul Christoffersen⁸,
- 4 Jemma Wadham⁹, Samuel H Doyle¹⁰, Bryn Hubbard¹⁰
- ¹School of Earth and Ocean Sciences, Cardiff University, Cardiff, UK (<u>prior-jonesm@cardiff.ac.uk;</u>
- 6 <u>bagshawe@cardiff.ac.uk</u>)
- 7 ² School of Engineering, Cardiff University, Cardiff, UK
- 8 ³ Department of Aerospace Engineering, University of Bristol, Bristol, UK
- ⁴ Laboratory of Hydraulics, Hydrology and Glaciology, ETH Zürich, Zürich, Switzerland
- 10 ⁵Geological Survey of Denmark and Greenland, Copenhagen, Denmark
- 11 ⁶ Niels Bohr Institute, University of Copenhagen, Copenhagen, Denmark
- 12 ⁷ Center for Earth Observation Science, University of Manitoba, Winnipeg, Canada
- 13 ⁸ Scott Polar Research Institute, University of Cambridge, Cambridge, UK
- 14 ⁹ School of Geographical Sciences, University of Bristol, Bristol, UK
- ¹⁰ Department of Geography and Earth Sciences, Aberystwyth University, Aberystwyth, UK

16 ABSTRACT

Subglacial hydrological systems require innovative technological solutions to access and observe. Wireless sensor platforms can be used to collect and return data, but their performance in deep and fast-moving ice requires quantification. We report experimental results from Cryoegg: a spherical

- 20 probe that can be deployed into a borehole or moulin and transit through the subglacial hydrological 21 system. The probe measures temperature, pressure and electrical conductivity in-situ and returns all
- 22 data wirelessly via a radio link. We demonstrate Cryoegg's utility in studying englacial channels and
- 23 moulins, including in-situ salt dilution gauging. Cryoegg uses very high frequency (VHF) radio to
- transmit data to a surface receiving array. We demonstrate transmission through up to 1.3 km of cold
- ice a significant improvement on the previous design. The wireless transmission uses Wireless M-Bus
- 26 on 169 MHz; we present a simple radio link budget model for its performance in cold ice and
- 27 experimentally confirm its validity. Cryoegg has also been tested successfully in temperate ice. The
- battery capacity should allow measurements to be made every two hours for more than a year. Future
- 29 iterations of the radio system will enable Cryoegg to transmit data through up to 2.5 km of ice.

30 INTRODUCTION

31 The presence and behaviour of liquid water in the subglacial environment governs the response of ice 32 to climate warming. Meltwater generated on the surface makes its way to the bed via networks of 33 moulins, cracks and crevasses (Chu, 2014; Flowers, 2018). Once at the bed, it flows to the ice margins 34 either through a subglacial drainage network consisting of inefficient linked cavities (Iken and 35 Bindschadler, 1986; Walder, 1986; Kamb, 1987), efficient channels carved into rock, ice or the 36 sediment below (Röthlisberger, 1972; Nye, 1976; Clarke, 1987; Ng, 2000), or a combination of both 37 (Hoffman and others, 2016). The configuration of the drainage network determines the subglacial 38 water pressure and how much of the ice-bed interface is in contact with liquid water. Contact 39 promotes sliding (Kamb, 1970; Iken, 1981; Schoof, 2010), which in turn can cause ice to accelerate 40 downstream. In recent years, the relationship between meltwater supply and ice acceleration has 41 been reevaluated in light of observations from the margins of the Greenland Ice Sheet that 42 demonstrate a seasonal evolution of subglacial drainage systems (Chandler and others, 2013; 43 Tedstone and others, 2015) commonly observed in Alpine systems (Nienow and others, 2005). Early 44 in the melt season, an increased flux of meltwater is routed to the bed and the low capacity, 45 inefficiently linked cavity system is forced to expand, forming efficient channels that can transport 46 substantial volumes of water. This reduces the area of the bed in contact with water, and potentially

regulates the flow of ice (Sole and others, 2011; Tedstone and others, 2015; Nienow and others, 2017;
Flowers, 2018). The defining feature of these different drainage configurations is the water pressure:
channelised systems operate at lower pressure than linked cavities, thus measurement of the
subglacial water pressure can be used to determine the likely structure of the drainage system, and
hence the acceleration response of the ice to increased surface melt inputs.

52 In addition to water pressure, other parameters may provide clues as to the structure of the drainage 53 system, but distinction between drainage system types is challenging. Temperature can be used to 54 assess whether the bed is at the pressure melting point, and the residence time of water in the system 55 can be used, in conjunction with pressure, to assess how efficiently the meltwater transits the system. 56 Long residence times are common in linked cavity systems, which results in prolonged contact 57 between meltwater and subglacial sediment (Tranter and others, 2002). This promotes chemical 58 weathering and changes the composition of the meltwater, so meltwater chemistry is a good indicator 59 of drainage system structure. Chemical composition is challenging to assess in situ, but a 60 measurement of the total dissolved solids can be easily obtained via a measurement of electrical 61 conductivity (Hubbard and others, 1995).

62 These three parameters (pressure (P), temperature (T) and electrical conductivity (EC)) are relatively 63 easy to measure via electronic sensors and can thus be combined to provide information on subglacial 64 drainage that could not be detected from the surface. Measuring these parameters subglacially is, 65 however, extremely challenging, particularly beneath thick, fast flowing ice. Yet it is these fast-flowing 66 sectors that govern the response of large ice masses to climate warming, since they transfer significant volumes of ice to the ocean (Pritchard and others, 2009; van den Broeke and others, 2016). Drilling 67 68 boreholes through a glacier to access subglacial hydrological channels is logistically demanding and 69 disturbs the system that is under study. Once boreholes are drilled, implanting cabled sensors is 70 possible, but ice motion causes cables to flex and eventually break (Iken and others, 1993; Doyle and 71 others, 2018). In fast-flowing ice, data capture is thus limited to days or weeks. An alternative method 72 is therefore required that can capture these relatively simple electrical measurements and return 73 them to the surface without requiring a physical connection. A wireless radio frequency (RF) system 74 is ideal and there is a long history of the use of RF propagation through ice (see Plewes and Hubbard, 75 2001 for a review).

76 Here, we present trials of Cryoegg, a wireless sensor platform for use in deep ice. The use of a radio 77 link for subglacial telemetry has been proven by the work of the Glacsweb programme (Martinez and 78 others, 2004; Hart and others, 2019) and the WiSe project (Smeets and others, 2012). Previous work 79 (Bagshaw and others, 2014) showed that a 'Cryoegg' concept was feasible, namely a spherical sensor 80 platform containing all sensor, radio and datalogger components that could fit in a standard borehole 81 and travel through subglacial meltwater pathways: the electronics could be made sufficiently 82 compact, and that the radio link worked through up to 500m of ice. However, the radio link design 83 chosen proved unsuitable for performance in very deep ice, so design improvements were required. 84 In this paper we describe the redesign of Cryoegg to give enhanced radio link performance and show 85 the outcomes of field trials at sites in Greenland and the Swiss Alps.

- In order to measure subglacial hydrological properties in deep polar ice, the enhanced Cryoegg had to
 meet or exceed the following engineering constraints:
- An outer diameter of 120 mm or less, to fit into a standard ice core borehole
- A radio link capable of reaching the surface through 2,500 m of ice, the mean bed depth in central Greenland (Morlighem and others, 2017)
- Survive and measure water pressure of up to 25 MPa (250 bar, equivalent to a water column of 2,500 m)
- Measure temperature, typically in the range from -30 to 0 °C
- Measure EC, typically in the range from 0.002 25 mS m $^{-1}$ (2 250 μS cm $^{-1})$
- A battery life capable of sustaining one measurement every 12 hours for a period of one year

96 RADIO LINK DESIGN

97 The success of the instrument depends principally on the performance of the radio link. The 2012 98 design (Bagshaw and others, 2014) used a simple frequency shift keying (FSK) transmitter operating 99 on 151 MHz and demonstrated a maximum range of 500 m in wet ice. To achieve a greater range, we 100 investigated alternative frequencies and transmission schemes. The power of a radio wave 101 propagating in "free space" (e.g. in air or vacuum) reduces according to an inverse square law with 102 distance - known as "geometric attenuation". When the propagating wave reaches a receiving 103 antenna, the ability of that antenna to extract power from the incoming wave is the "effective 104 aperture", and this depends upon the wavelength of the incoming wave. Antenna performance is 105 more usually characterised using the antenna gain, which is the ratio of the antenna's effective 106 aperture in the direction of the main beam to the effective aperture of an "ideal" isotropic antenna 107 that receives signals equally well in all directions.

These effects are collectively described by the free space path loss equation, sometimes known as the Friis transmission equation, which describes how a radio link performs in free space. The equation assumes that the antennas are optimally pointed at one another and that their polarisations match, otherwise there are further losses associated with pointing error and polarisation mismatch. The original paper (Friis, 1946) presents the equation in terms of effective aperture, and in linear units. The more commonly used version quoted here is expressed in terms of antenna gain and uses decibel units.

115 Equation (1), adapted from Griffiths (1987, p. 12), is the free space path loss equation in decibel units:

116
$$P_{rx} = P_{tx} + G_{tx} + G_{rx} - 20 \log_{10} \left(\frac{4\pi d}{\lambda}\right)$$
 (1)

117 P_{rx} is power at the receiver, in dBW (dB relative to 1 watt) 118 • P_{tx} is power transmitted by the transmitter, in dBW 119 • G_{tx} is the gain of the transmitting antenna, in dBi (dB relative to the performance of an isotropic antenna) 120 • *G*_{rx} is the gain of the receiving antenna, in dBi 121 122 d is the distance between the transmitting and receiving antennas in metres • 123 λ is the wavelength of the transmission. •

124 The last term of Equation (1) is known as the "free space path loss" (FSPL) and combines the geometric 125 attenuation due to distance with the apparent wavelength-related attenuation caused by the effective 126 aperture of the antennas. Consequently, the free space path loss equation gives us the rule of thumb

- 127 that lower frequencies (longer wavelengths) appear to propagate further than higher frequencies.
- 128 FSPL over 2500 m ranges from 70 dB at 30 MHz to 99 dB at 868 MHz (Table 1), depending on frequency.

Table 1: Values of free space path loss in dB for several frequencies used by previous subglacial wireless transmission systems, the industrial standard LoRaWAN and the redesign of Cryoegg (see results section).

System	WiSe ^a	eTracer ^b , Cryoegg (2012) ^b , Glacsweb (2012 – present) ^c	Glacsweb (2004 –2006) ^d	LoRaWAN ^e	Cryoegg (2019)
Frequency (MHz)	30	150	433	868	169
Distance (m)					
100	42.0	56.0	65.2	71.2	57.0
500	56.0	70.0	79.2	85.2	71.0
1000	62.0	76.0	85.2	91.2	77.0
1500	65.5	79.5	88.7	94.7	80.5
2000	68.0	82.0	91.2	97.2	83.0
2500	70.0	83.9	93.1	99.2	85.0

a) Smeets and others, 2012; b) Bagshaw and others, 2014; c) Hart and others, 2019 d) Martinez and
others, 2004; e) https://lora-alliance.org/about-lorawan/

134 FSPL applies to all radio links regardless of the propagating medium. However, where the medium is 135 lossy the signal is further attenuated by the interaction between the wave and the medium. Ice is one such lossy medium. Whilst simple models can predict RF attenuation in pure ice, in reality, glacier ice 136 137 is heterogeneous, varying in temperature, pore water and impurity content and it contains cracks, 138 water pockets and debris. Ultra-high frequencies (UHF, 300 MHz-3 GHz) have been effective for 139 transmission through deep, cold and uniform ice (Lewis and others, 2015), but any presence of water 140 in this matrix quickly reduces success due to scattering and attenuative losses. The high frequency 141 (HF, 3-30 MHz) and very high frequency (VHF, 30-300 MHz) bands have good penetration through ice, with wavelengths longer than typical englacial water bodies encountered along the transmission path 142 143 (asserted by Smeets and others, 2012).

144 The attenuation of electromagnetic waves in glacial ice is reported in the study of high-energy 145 neutrinos (Barwick and others, 2005; Barrella and others, 2011) as *attenuation length* (L_a) in metres. 146 This is related to *attenuation coefficient* (*a*) in decibels per metre by Equation (2) (Barrella and others, 147 2011):

148
$$L_{\alpha} = \frac{1}{\ln \sqrt{10^{\frac{\alpha}{10}}}}$$
 (2)

We can rearrange and simplify Equation (2) to convert attenuation length to attenuation coefficient
 in dB m⁻¹ (Equation (3)). Some typical values of attenuation length are shown converted to dB per
 kilometre in Table 2.

$$152 \qquad \alpha = \frac{20}{(\ln 10)L_{\alpha}} \tag{3}$$

Attenuation length (L _α), metres	Attenuation coefficient (α), dB km ⁻¹		
100	86.9		
200	43.4		
300	29.0		
400	21.7		
500	17.4		
600	14.5		
700	12.4		
800	10.9		
900	9.7		
1000	8.7		
1500	5.8		
2000	4.3		
5000	1.7		

154 Table 2: Attenuation length and corresponding attenuation coefficient from 100 to 5000 m

156 Attenuation lengths determined experimentally vary slightly by frequency (Barwick and others, 2005), 157 with lower frequencies generally having longer attenuation lengths (and hence lower attenuation 158 coefficients). Mitigating against both free-space and ice-related losses therefore points towards the 159 use of lower frequencies for radio links within ice. However, low frequencies imply long wavelengths, 160 which in turn requires physically large antennas, as an efficient antenna needs to be at least ¼ of a wavelength long. The WiSe system at 30 MHz has a wavelength of 10 metres and used half-wavelength 161 162 dipole receiving antennas which were five metres long (Smeets and others, 2012). Large antennas 163 become impractical to work with in the field, and we had the additional challenge of needing to fit our 164 transmitting antenna into the 120 mm diameter Cryoegg enclosure. Our previous work used 151 MHz 165 (wavelength 2 m), which had given satisfactory performance and allowed the receiving antenna to be 166 easily carried in the field. The very small size of the Cryoegg enclosure meant that there was limited 167 value in going to higher frequencies as the benefits of having a better-matched transmitting antenna 168 were far outweighed by the additional free-space and ice-related losses. We therefore looked for a 169 system that could operate in the VHF band (30-300 MHz).

For frequencies in the VHF range (30-300 MHz) the attenuation lengths reported (Barwick and others,
2005; Barrella and others, 2011) range from 200 m to 3000 m. We take the worst-case figure of 200

172 m (43.4 dB km⁻¹) as our design criterion for working in warm, wet ice and 400 m (21.7dB km⁻¹) as a

173 conservative estimate for cold, dry ice (Table 2).

175 METHODS

176 Choice of transmission scheme


177 The term "transmission scheme" encompasses all the technical aspects of the radio link - the modulation, error corrective coding, packetisation, and higher-level protocols. We required a 178 179 commercially available system designed for long battery life and for sending small amounts of data 180 over long distances, often sold as low power wide area networks (LPWAN) or "internet of things" (IoT) systems. We selected Wireless M-Bus mode N1 (European Committee for Standardisation, 2013), 181 182 which is intended for use in utility metering. It is designed to offer very long battery life and sends 183 data at 2.4 kbit s⁻¹. It incorporates error corrective coding, which ensures that data received over the 184 link does not contain errors introduced in transit, and has optional cryptographic protection for 185 security. It operates on the 169 MHz frequency band, which is available for general license-free use in 186 Europe (CEPT ECC Recommendation 70-03). Wireless M-Bus is an open standard and a number of manufacturers provide implementations of it. This gives confidence that the technology will remain 187 188 available, whereas a proprietary system carries a risk of the product being discontinued.

The Radiocrafts RC1701HP-MBUS4 modem used provides 0.5 W (27 dBm) power output on the 169 MHz band. One module is fitted to the printed circuit board (PCB) inside Cryoegg, configured as a transmitter. For the receiver, we use Radiocrafts RC1701HP-MBUS4 demo kit with one board configured as a receiver. This board connects to a PC and outputs the decoded data received over the radio link. A Python script running on the PC applies a timestamp to the received packet data and records it in a log file. This approach avoids the need to have a source of accurate time on board Cryoegg.

196 Antenna selection

The transmitting antenna inside Cryoegg is constrained by the physical size of the spherical case. We chose the HA.10 from Taoglas which is designed for the 169 MHz band, consisting of a pair of 30mm long air-cored helical elements and a matching network. To minimise use of conductive materials around the antenna, the upper hemisphere is devoted to the antenna and the remaining electronics fit into the lower hemisphere (Fig. 1a).

202



203

Fig. 1 – a) Cryoegg with upper casework removed b) receiving antenna mounted on plastic frame

The receiving antenna on the surface provides additional gain to the system to help overcome the attenuation through the ice, and to compensate for the small size of the transmitting antenna – an ideal antenna would be 450 mm long, yet Cryoegg's diameter is only 120 mm. We elected to use a

- 208 pair of crossed Yagi-Uda antennas (Innovantennas), which provide a gain of around 8 dB individually.
- They are combined through a 90° hybrid combiner (Mini-Circuits part number ZMSCQ-2-180BR+)
- which makes them behave as a single circularly polarised antenna, but at the expense of 3dB loss in
- the combiner. By transmitting with linear polarisation and receiving with circular polarisation, we
- 212 make the radio link performance relatively independent of Cryoegg's orientation. This technique was
- also used by the WiSe project team (Smeets and others, 2012).
- The receiving antenna is one metre long. In order to minimise disturbance to the antenna pattern caused by metal parts close to the antenna elements, we used a modular plastic frame (Quadro) to support the antenna (Fig. 1b).
- 217 The rediction methods of Churches according a several DE test show her (at the
 - The radiation pattern of Cryoegg was assessed in a screened RF test chamber (at the Wolfson Centre for Magnetics in Cardiff) lined with absorbent ferrite material to prevent multipath. A log-periodic antenna was used to receive the signal and the signal level was observed using a spectrum analyser in peak hold mode.
 - 221 Link budget
 - A link budget is used to evaluate whether an attenuation-limited radio link will work in practice. Starting with the power output of the transmitter, gains and losses in the system are totalled up and
- compared to the sensitivity of the receiver. To allow some margin for unexpected attenuation, we aim for a received power level several dB higher than the sensitivity. Link budgets are traditionally calculated in decibel units as this allows the gains and losses to be added and subtracted (rather than multiplied and divided). Hence we use decibel units of power, such as dBW: decibels relative to one watt, (0 dBW = 1W) or dBm: decibels relative to one milliwatt (0 dBm = 1mW = -30 dBW, and +30 dBm
- 229 = 1 W = 0 dBW).

The link budget calculation (Table 3) assumes a 2000 m borehole through cold ice, with the attenuation coefficient estimated at 21 dB km⁻¹. The performance of the transmit antenna was relatively poor, and so we estimated its gain at -15 dBi (dB relative to an isotropic antenna) based on data from the manufacturer. For the 2000 m example shown here, the received signal margin is 10.5 dB.

Table 3: Link budget calculation for Cryoegg in 2000 m borehole in cold ice (gains are positive values,

237 losses are negative).

238

			Link budget contribution	Units
Transmitter power	0.5	W	27.0	dBm
Coupling loss			-0.5	dB
Transmit antenna gain			-15.0	dBi
Distance to receiver	2	km		
Frequency	169	MHz		
FSPL			-83.0	dB
Attenuation coefficient for cold ice	21	dB/km		
Ice related loss			-42.0	dB
Crosspolarisation loss			-3.0	dB
Receive antenna gain			8.0	dBi
Total power at receiver			-108.5	dBm
Receiver sensitivity	-119	dBm		
Margin			10.5	dB

239 Sensors

240 The Keller PA-20D pressure sensor (max. 250 bar) has a vacuum-sealed membrane and communicates 241 with the microcontroller via the digital I²C interface (Inter-Integrated Circuit; (UM10204 I2C-bus 242 specification and user manual, 2014)). It provides internal temperature compensation, and supplies a 243 temperature reading alongside the pressure reading, although the manufacturer does not guarantee 244 its performance at temperatures below 0 °C. Hence we provided our own independent temperature 245 sensor (details below). The sensor provides a 16-bit pressure reading to the microcontroller but uses 246 only half the available range (the rest being used to allow it to report pressures slightly beyond the 247 calibrated range). This means that the smallest pressure step reportable is 7.6 millibars. The nominal 248 total error band is 1% of full scale, i.e. 2.5 bar, but in practice we found we could reliably record 249 changes in water pressure down to 0.1 bar (1m hydrostatic pressure) during field experiments.


250 The temperature and EC sensors are adapted from earlier designs (Bagshaw and others, 2012, 2014). 251 The EC sensor consists of a square wave oscillator which supplies a 500 kHz waveform to a potential 252 divider consisting of a precision resistor and a pair of sense electrodes. The sense electrodes are a pair 253 of M3 stainless steel hex-headed bolts that protrude through the case. The AC waveform from the 254 midpoint of the potential divider passes through a precision rectifier and RC filter to produce a DC 255 voltage that varies inversely with EC between the sense electrodes. This is sampled by the 256 microcontroller's analogue-to-digital converter (ADC) and the resulting digital value is reported over 257 the radio link. The temperature sensor is a Pt1000 platinum resistance device, used in a full-bridge 258 configuration with three fixed resistors. It is driven by a current source and measured using an 259 instrumentation amplifier, with the output fed to the microcontroller's ADC. Cryoegg reports the 260 digital value from the ADC over the radio link, allowing calibration to be carried out externally. The 261 Pt1000 resistor is mounted to the back of one of the EC sense electrodes with a thermal pad, ensuring that it has thermal but not electrical contact. 262

263 Microcontroller selection

- 264 To maximise battery life, Cryoegg has a "sleep" mode where most subsystems were powered down between measurements. The STM32L433RCT6P microcontroller (STMicroelectronics) has a built-in 265 Real Time Clock module (RTC) which uses an external quartz crystal oscillator to provide reliable 266 timekeeping at very low power. This "wakes" the microcontroller when it is time for the next 267 268 measurement. The RTC draws around 500 nA at 3.3 V with the rest of the microcontroller shut down. 269 The microcontroller has a 32-bit ARM Cortex M4 processor that can be clocked at up to 80 MHz, 256 270 kB of flash memory and 64 kB of RAM, and includes the ADC for the temperature and EC sensors; I²C 271 interface for the pressure sensor; and Universal Asynchronous Receiver/Transmitter (UART) for 272 communicating with the radio module. It also controls several power switches that enable and disable 273 power to other parts of the circuit.

274 Power supply design and power consumption

- 275 The radio modem module has relatively high power consumption during transmit requiring 500 mA
- at 3.3 V for less than 500 ms during each transmission, which puts a lot of demand on the battery and
- power supply to be able to supply this peak current. A lithium-polymer rechargeable pouch cell can
- supply sufficient peak current and be recharged between tests. The battery selected has a 3.7 V
- 279 nominal voltage and a capacity of 400 mAh.
- A regulator IC with an enable input (ON Semiconductor NCP115ASN330T2G) supplies 3.3 V to the sensors when enabled, and also provides the ADC reference voltage to the microcontroller. Other parts of the circuit are supplied with unregulated battery voltage.
- To estimate the battery life, we measured the power consumption of Cryoegg during transmission and during sleep mode in the lab, using a logging multimeter (Mooshim Engineering Mooshimeter) that could measure voltage and current from the battery simultaneously. The measure-and-transmit cycle takes 3.2 seconds and consumes 0.5 J. The sleep mode current consumption proved to be too low for the meter to measure (the lowest current it can record is 5 μA). We therefore assume that the sleep mode current consumption is that of the microcontroller only (since everything else is disabled) and
- take the value quoted in the microcontroller datasheet of 500 nA.



291 Fig. 2 - Cryoegg projected battery life with varying sampling rate

290

The projected battery life based on these measurements is over 6 years at two measurements per day. Even allowing for some self-discharge in the battery, this gives scope to increase the measurement frequency. A measurement every 2 hours (i.e. 12 times per day) gives a battery life of just over a year. Battery life estimates at different sampling rates are shown in Fig. 2. Table A1 in Appendix A shows an example battery life calculation in more detail.

298 Mechanical design

299 We aimed to provide a simple and robust mechanical design that was straightforward to assemble for 300 testing. The spherical casework is machined in two halves from acetal copolymer, a hard engineering 301 plastic (Fig. 1a). The sensor PCB sits directly onto the bottom of the cylindrical internal void and is 302 secured in place by two M3 threaded spacers, then potted with Raytech Liquid Rubber. There are two 303 further PCBs that mount above the sensor PCB, which interconnect using multiway connectors. The 304 processor PCB contains the microcontroller and associated components, and also provides mechanical 305 support for the battery. The radio PCB is uppermost and supports the radio module, antenna 306 connector, battery connector and headers for programming and debugging. The pouch cell battery is 307 sandwiched in the gap between the radio and processor PCBs. The antenna PCB connects to the radio 308 PCB via an SMA connector and is supported by a groove in the crown of the upper hemisphere. This 309 design allows the upper hemisphere to be easily removed for access to the electronics, and to connect 310 the battery before deployment. The upper and lower hemisphere seal with a rubber O-ring and are 311 held in place by eight machine screws.

313 Software

The software on the Cryoegg microcontroller is written in C, using the STM32 Hardware Abstraction Libraries. The software goes through the following steps:

- Power up the sensors
- Make measurements
- Power down the sensors
- Power up the radio module
- Pack the sensor data into a data packet
- Send the data packet to the radio module to be transmitted
- Power down the radio module
- Set the sleep timer for the next measurement
- Enter deep sleep ("SHUTDOWN") mode

325 On awakening from SHUTDOWN mode, the program restarts from the beginning, thus giving an

326 endless loop. The measurements are transmitted immediately after being made, and no data is stored

- 327 on Cryoegg, since we do not expect to retrieve Cryoegg after deployment.
- 328 Field testing

329 We conducted three field tests during July and August 2019 at two sites in Greenland (EastGRIP drill

330 site and Sermeq Kujalleq/Store Glacier) and one in Switzerland (Rhône Glacier). The objectives of the

field tests were to verify the operation of Cryoegg and its receiver in a real glacial environment. We

332 wanted to demonstrate the operation of the radio link in both cold deep ice and shallower temperate

ice, and confirm the mechanical integrity of the device when subjected to hydrostatic pressure in fluid-

filled boreholes.




Fig. 3 - Velocity map of Greenland ice sheet flow, showing locations of EastGRIP and Sermeq Kujalleq
 (Store Glacier) test sites in fast flowing ice. Data from MEaSUREs dataset of annual Greenland

velocity for 2018 (Joughin and others, 2010; Joughin, 2017).

338 East Greenland Ice Core Project site (EastGRIP)

EastGRIP is located at N75° 38.05' W036° 00.22' on the North East Greenland Ice Stream (NEGIS), the 339 largest ice stream in Greenland, which drains 340,000 km² of the ice sheet and extends for over 1000 340 341 km inland (Fig. 3). Approximately 150 km from the onset, it reaches speeds of 65 m a⁻¹ (Joughin and others, 2010; Karlsson and Dahl-Jensen, 2015). The East Greenland Ice core Project (EastGRIP) is a 342 unique project drilling an ice core into 2.5 km of fast flowing ice to investigate ice stream beds 343 344 (www.eastgrip.org). In summer 2019, the core had been drilled to 2 km depth, leaving behind a 2 km 345 borehole filled almost completely with drill fluid. The purpose of our field trial at this site was to obtain 346 a range test for the radio link and a pressure test for the mechanical design. Cryoegg was deployed in 347 a mesh bag and attached to the main winch (Fig. 4). The orientation of Cryoegg within the bag was not fully controlled due to its spherical shape, although the pressure sensor protruding through the 348 349 mesh helped to prevent it from shifting. The receiving antenna was set up in the drill trench, close to 350 the winch (Fig. 1b). Cryoegg was lowered and raised into the borehole several times and the received 351 signal strength (RSSI) and live data stream monitored at the surface, adjacent to the top of the 352 borehole.



Fig. 4 – Cryoegg ready for deployment on the EastGRIP ice core winch.

355

356 *RESPONDER site at Sermeq Kujalleq (Store Glacier)*

357 We tested Cryoegg at an inland site (N70° 33.889' W50° 04.558') at Sermeg Kujalleg (Store Glacier), 358 the third fastest outlet glacier in West Greenland (Fig. 3). It has a catchment of 35,000 km² that 359 includes supraglacial lakes that periodically drain via cracks and moulins, several on an annual basis 360 (Chudley and others, 2019). The glacier experiences changes in ice flow associated with sudden 361 injections of meltwater to the pressurised drainage system (Doyle and others, 2018), but the link 362 between surface lake drainage and the subglacial hydrology is poorly defined, primarily because instrumenting a draining lake with cabled sensors is near-impossible. There is extensive supporting 363 364 data available on the subglacial bed structure, lake drainage frequency and ice strain rates (Hofstede 365 and others, 2018; Young and others, 2018; Chudley and others, 2019). Sermeq Kujalleq (Store Glacier) is the site of the RESPONDER project ("Resolving Subglacial Properties, hydrological networks and 366 367 dynamic evolution of ice flow on the Greenland Ice Sheet", https://www.erc-responder.eu/), offering 368 access to the glacier bed through hot water drilling. The ice is approximately 1 km thick at this site 369 (Morlighem and others, 2017) and moving at 600 m a⁻¹ (Chudley and others, 2019). Bed access holes 370 were hot water drilled in July 2019.

- A surface propagation test assessed the range of data transmission through air by monitoring the RSSI
- and live data stream as the receiving antenna was deployed at a fixed site and Cryoegg hand-carried
- over a distance of 1.6 km. A hand-held GPS receiver was used to record the position of Cryoegg as it
- was carried, and the fixed position of the Cryoegg receiver.





378

The Rhône Glacier is located at N46°34.32′ E8°22.58′ in the Swiss Alps and is one of the most studied glaciers, with records of front position dating back to the 17th Century (Church and others, 2019). The 16 km² glacier is at the pressure melting point throughout and there is an active subglacial drainage network. The glacier is the focus of an intensive subglacial monitoring project, with artificial moulins drilled via hot water in 2018. The moulins remained active in August 2019, when we deployed Cryoegg on the end of a rope tether.


Salt dilution gauging (Moore, 2005) was used to estimate moulin discharge. A known quantity of tracer, sodium chloride (NaCl, "table salt") was added to the supraglacial stream approximately 25 m upstream of the moulin. The discharge can then be calculated from EC readings and the concentration of NaCl added. EC was measured in the supraglacial stream by a Keller DCX-22-CTD 15 m upstream of the moulin and also recorded and transmitted once per second by a Cryoegg lowered into water at the bottom of the moulin.

392 **RESULTS**

393 Laboratory RF tests

394


- 395 RF power meter assessments confirmed the assumption (Table 3) that the transmitter put out the full
- 396 +27 dBm (0.5 W) during each transmission.




397

398 Fig. 6 – Horizontal radiation pattern of Cryoegg measured in the RF test chamber

The horizontal radiation pattern was measured at eight points around the circumference and is shown in Fig. 6. For this measurement the receiving antenna (a log-periodic) was vertically polarised (a brief check showed that this gave a larger signal than when horizontally polarised) and Cryoegg was orientated with its case split line horizontal. Two complete revolutions were measured to check consistency. The pattern is largely omnidirectional, with <2 dB of variation between all the measurements.






To simulate the RF performance in the borehole, we re-oriented the Cryoegg to have the same orientation as it would have in the borehole, with the split line vertical and normal to the receiving antenna boresight – i.e. with the crown of the upper half pointing towards the receiving antenna, and the sensor ports pointing away. The receiving antenna was vertically polarised. The results in Fig. 7 show that the signal level is significantly lower (10-12 dB) than in the horizontal plane, and that the pattern is not omnidirectional; there is a 6 dB variation as the unit is rotated.

413

415 Surface propagation at Sermeq Kujalleq (Store Glacier)



416

Fig. 8 – a) RSSI recorded by receiver during surface range test at Sermeq Kujalleq (Store Glacier) and
modelled received signal strength.). b) shows the ground elevation (height above the WGS84 ellipsoid)
along the route taken. The black vertical line in both plots shows the point where the transmitter went
beyond the line of sight to the receiver due to the ice sheet's surface topography.


421 Fig. 8a shows recorded signal strength for successfully received data packets against range from the 422 receiver. The terrain profile in Fig. 8b was produced from ArcticDEM v3 423 (https://doi.org/10.7910/DVN/OHHUKH) 2 m mosaic values extracted to match the GPS positions recorded in the field. A line-of-sight binary was calculated using the QGIS visibility analysis plugin with 424 425 the receiver height set at 1.5 m and the transmitter height at 1 m.

426 Deployment at Sermeq Kujalleq (Store Glacier)

427 Cryoegg was lowered into a hot water drilled borehole and a moulin. One deployment attempt was 428 made in the hot water drilled borehole, but the borehole proved too narrow for Cryoegg to pass 429 through. One data point was obtained with Cryoegg in the borehole about 400 m below the surface, 430 but it was impossible to proceed further because of borehole refreezing. Moulin deployment was attempted in a very large moulin (measured at 4.3 m³s⁻¹ discharge at the time of deployment) adjacent 431 432 to the drill site. Cryoegg was caught in a series of plunge pools and eventually the force of the water caused it to break free from its tether and it was rapidly swept away out of range. We only obtained 433 434 a few data points before losing the signal.

435 Downhole propagation at EastGRIP

At EastGRIP, the borehole is filled with ESTISOL 240 drill fluid rather than water (Sheldon and others, 436 437 2014). Previous tests at the site (Bagshaw and others, 2018) demonstrated that the fluid had minimal 438 impact on signal propagation. Fig. 9 shows the RSSI plotted against depth. Depth is linearly 439 interpolated between depth-measured winch halt points (shown as vertical gridlines on Fig. 9), which 440 is a fair assumption because the winch motor speed was constant between these halts. The firmware 441 was configured to produce a burst of 16 packets, one per second, and then wait for 60 seconds before 442 the next burst. This accounts for the clustered data points on Fig. 9, as all the successfully received 443 packets are plotted. There are large variations (>10 dB) in signal level at 300, 400 and 500 m, coinciding 444 with the point where the winch was halted, and even retrograde paths, for example between 400 and 445 500 m; 700-850 m; 1000-1100 m. The deepest point at which packets were successfully received was 446 1340m below the surface. No packets were received as Cryoegg was pulled back up to the surface. 447 When Cryoegg was retrieved, it was found to be full of drill fluid, indicating that it had leaked under 448 pressure.




449


450 Fig. 9 – Received signal strength of successfully-received data packets during a test in the EastGRIP

borehole, together with modelled signal strength (explained in "radioglaciological implications"

452 section)








456 The hydrostatic pressure increased with depth (Fig. 10) – the small offset is because the drill fluid is 457 around 50 m below the surface level. Data was obtained down to 1340 m, although significant packet 458 loss occurred beyond 1250 m. The two temperature plots represent data from the two different 459 temperature sensors. The Pt1000 sensor is inside the case, hence it records a higher temperature for 460 a given depth when compared with the Keller, which has direct contact with the drill fluid. The 461 conversion equation from the Pt1000 reported value (which is in arbitrary units) to temperature is 462 based on theory and has not been confirmed by calibration. Since the primary test objective was to test the radio link, we did not prioritise calibrating the temperature sensors prior to the field 463 464 deployment. The high degree of clustering of the data points from the Keller sensor suggests that it 465 contains some internal averaging, although this may be caused by the significant mass of the sensor body itself. Cryoegg was at ambient ice surface temperature before entering the borehole, so the 466 467 majority of the temperature data recorded here is simply the instrument cooling down to the ambient 468 englacial temperature. The spread in Pt1000 temperature results beyond 400m depth is caused by 469 self-heating of the sensor due to repeated measurements in quick succession, which becomes 470 apparent once Cryoegg has cooled to the ambient englacial temperature.

472 Moulin drainage at Rhône Glacier

473 Cryoegg did not reach the bed of the Rhône Glacier, 200m below the surface, via the moulin and instead appeared to be in a deep plunge pool 150m below the surface. Pressure readings received 474 475 from Cryoegg in real time confirmed that it was in up to 25m deep water. Fig. 11 shows the pressure 476 recorded whilst Cryoegg was in the plunge pool, and the corresponding received signal strength. There 477 was a sharp rise in pressure as Cryoegg was lowered into the water at 13:12:00 UTC and then a gradual 478 decline over the next 40 minutes. The gap in the data centred on 13:55 was an interruption in the data 479 logging. After the logging resumed, the water pressure had fallen to atmospheric pressure. The 480 reduction in water pressure coincided with a 10dB increase in received signal strength over the same 481 period.



482

Fig. 11 – Pressure and received signal strength from Cryoegg in a moulin on Rhône Glacier, 15th
 August 2019


485

487 Salt discharge gauging at Rhône Glacier

The salt discharge gauging experiment at Rhône Glacier consisted of injecting a 100 g l⁻¹ NaCl salt solution into a supraglacial stream, 25m upstream from a moulin. A Keller DCX-22-CTD in the supraglacial stream, 10m downstream from the injection point, measured the EC in the stream. Simultaneously, a Cryoegg in the moulin plunge pool (150m below the ice surface) measured the EC

and transmitted the measurements to the surface in real time.

493 Fig. 12 shows the results from the Keller logger in the stream alongside the results from Cryoegg in494 the moulin.



495

The discharge of the supraglacial stream was calculated by the salt dilution as 104 litres s⁻¹ (Moore, 2005), and the discharge within the moulin was slightly higher at 113 litres s⁻¹. The time between the two peaks was 60 seconds. The velocity of the water between the two instruments was 2.75 m s⁻¹ based on the transit time and the distance between them (15 m in the stream + 150 m down the moulin = 165 m).

 ⁴⁹⁶ Fig. 12 – comparison of salt wave passing Keller logger in the supraglacial stream with Cryoegg
 497 within the moulin. 15th August 2019.

504 **DISCUSSION**

505 Radioglaciological implications

506 Surface range test

507 The main objective of field testing was to verify the performance of Cryoegg in a real glacial 508 environment. The surface range test at Sermeq Kujalleq/Store Glacier (Fig. 8) confirmed that the 509 radiated output of Cryoegg matched our design calculations. We modelled the expected RSSI using 510 the conventional two-ray ground-reflection model (Bullington, 1947) used for VHF propagation. The 511 transmitter and receiver heights required were modelled using the DEM profile shown in Fig. 8b. The 512 transmitter and receiver parameters in the model were those from the link budget in Table 3. The 513 model produces a good fit to the real data, confirming that the parameters were indeed realistic 514 estimates. The variations in the signal strength data are because the DEM profile is derived from data 515 with a 2m resolution of the glacier surface terrain - but the glacier surface had undulations of around 516 0.5-1m in height. This caused variations in antenna height during the test that are not reflected in the 517 DEM profile. The later part of the test was beyond the line of sight to the receiver (as shown by the 518 elevation profile in Fig. 8b), which accounts for the step reduction in signal strength beyond 1100 m.

519 EastGRIP borehole test

520 The test in the EastGRIP borehole was intended to verify both the mechanical and electrical 521 performance of Cryoegg in a simulated deployment, and demonstrate the radio link through deep 522 glacial ice. No data was received beyond 1340 m depth and on return to the surface Cryoegg was 523 found to have failed and stopped transmitting. This appears to have been a mechanical failure. Drill 524 fluid was found in the Cryoegg housing after the tests. The fluid is non-conductive and so should not 525 have caused any electrical damage, but the hydrostatic pressure is likely to have affected some 526 components. The most likely failure points are the battery (which being a soft "pouch cell" type has 527 no protection from pressure) and its connectors, which may be forced apart by non-conductive fluid 528 under high pressure, breaking the circuit.

529 An important question remains: if the leak had not occurred and Cryoegg had continued operating 530 down to the bottom of the borehole (2000 m down), would we have received data at that depth? To 531 determine this, we estimate the RF performance based on the recorded data.

532 The receiver sensitivity (the minimum decodable signal strength) is -119 dBm (1.25 fW), and in other 533 range tests (Fig. 8) we succeeded in decoding signals down to this level (RC1701xx-MBUS Datasheet, 534 2018). However, the RSSI data for the EastGRIP borehole (Fig. 9) show that the weakest signals 535 received were at -103 dBm, 16 dB above the minimum receivable level. This suggests the system would 536 have continued working beyond 1340 metres depth without the mechanical failure. To forecast signal 537 levels at greater depth, we modelled the expected RSSI in the borehole. The model is based on the 538 same method as for the link budget in Table 3, but with ice attenuation calculated based on the 539 glacier's internal temperature at each depth rather than assuming a constant value at all depths.

540 The temperature profile of the EastGRIP borehole was measured in July 2019 using the University of 541 Copenhagen's automated logger (Gundrestrup and others, 1994). We applied the temperature 542 measurements to an attenuation model (MacGregor and others, 2007, 2015) to give an estimate for 543 the ice-related attenuation down to 2100m below the surface (Fig. 9). To continue the temperature 544 profile to the glacier bed depth (2500m) we conjectured that the temperature would rise rapidly to 545 reach 0 °C at the bed, by analogy with the temperature profile at NorthGRIP (Dahl-Jensen and others, 546 2003). This represents a worst-case since the model produces very high attenuation (69 dB km⁻¹) at 547 0 °C. This produces a profile with relatively low attenuation (~7dB km⁻¹) in the coldest part of the 548 glacier, increasing rapidly at the surface and bed where the ice is warmer.

549 When we initially developed the model, we found that using the value of -15dBi for the transmit 550 antenna gain (which produced a good fit to the surface range test data) overestimated the RSSI in the 551 borehole, even at points very close to the surface. This suggests that the radiated signal from Cryoegg 552 was much lower in the borehole when compared to the surface range test. It is likely that the antenna 553 performance was affected by the substantial metal apparatus on the winch cable just above where 554 Cryoegg was attached. The dielectric constant of the drill fluid is not the same as in air and this may 555 also have affected the antenna matching. Consequently, we reduced the transmit antenna gain by 556 20dB (to -35dBi) in the model to produce a result that more closely fits the observed data. The final 557 model output with the reduced transmit gain is plotted in Fig. 9.

558 The modelled signal strength is higher than the real data at most depths (Fig. 9), although it matches 559 the peak between 800 and 900 m, and there are two further effects that help explain why.

560 Firstly, the large (>10 dB) variation in signal strength observed at several winch halt points is most 561 likely caused by Cryoegg rotating on the vertical axis as the wire rope twists. Cryoegg was oriented 562 with its split line horizontal during these tests, meaning that the transmitting antenna is oriented in an end-fire mode towards the receiver. The deceleration of the winch will result in some of the 563 564 momentum of Cryoegg and the cable being converted into torsional forces on the winch cable, with 565 Cryoegg twisting back and forth on the end of the cable. This will affect the radiation from the antenna 566 because the radiation pattern seen by the receiving antenna is not uniform (Fig. 7). It is plausible that 567 the retrograde slope of the signal strength between 400 and 500m is caused by Cryoegg slowly 568 rotating on the winch cable, given the signal variation is comparable to that seen during the 400m and 569 500m winch halts.

570 Secondly, birefringence likely impacted the signal. Birefringence is a phenomenon exhibited by many 571 crystalline materials, where the refractive index varies depending on the polarisation and direction of 572 propagation of electromagnetic waves. Glacier ice affects the polarisation of VHF radio waves through 573 birefringence, and radar studies use this effect to infer details about the crystal fabric of the glacier 574 ice (Hargreaves, 1978; Li and others, 2018). Birefringence splits the wave from Cryoegg into two 575 elliptically-polarised waves of opposing chirality with differing phase velocities. The relative phase 576 delay between the two waves appears at the receiver as an apparent change in polarisation, and 577 therefore it is possible that the wave could arrive at the receiving antenna on the opposite polarisation to the antenna, resulting in a "null" - a significant reduction in received signal strength. Since this 578 579 phase delay relates to distance travelled through the medium, this effect would produce nulls at 580 specific depths corresponding to relative phase delays of 180 degrees. The signal strength plot (Fig. 9) appears to show a number of nulls-notably at around 600 m and 1040 m, which are most likely caused 581 582 by birefringence. Multipath reflection effects, such as when the signal reflected off a surface interferes destructively with the direct signal at the receiver (Griffiths, 1987, 102–104), are the other possible 583 584 explanation, but these can be discounted because there is no obvious candidate for the reflecting 585 surface. The geometry required to produce widely-spaced large nulls rules out horizontal reflectors 586 like the glacier bed or internal layers, and the shear margin is too far away (5 km) to produce this type of null. 587

The axial rotation of Cryoegg and the birefringent effects may also act in concert, which would explain why the signal variations are so large at some of the winch halt points (>10 dB at most locations). These two effects explain why the signal level drops below the modelled values: the model assumes constant antenna gain and matched polarisation, whereas in the real data the orientation varies

- (altering the transmit antenna gain) and the birefringence means that a proportion of the signal poweris transferred to the opposite polarisation and is lost to the receiver.
- 594 The birefringent effect could be mitigated against in future development by feeding the output of 595 each of the two crossed receiving antennas into a two-channel diversity receiver, which would then 596 be able to decode the signal regardless of its polarisation. This polarisation diversity technique has 597 previously been demonstrated mitigating polarisation nulls in HF ionospheric radio links (Stott, 2005) 598 where magnetoionic effects produce polarisation changes which are analogous to those produced by 599 birefringence (Davies, 1990).
- 600 Rhône glacier moulin test
- The performance of the radio link in the Rhône glacier moulin (Fig. 11) was satisfactory. We anticipated that the temperate ice and presence of flowing water would increase the overall attenuation. Fig. 11 shows that the signal propagating through 25 m of meltwater and a further 125 m of temperate ice to the glacier surface was attenuated to -90 dBm. This compares with the EastGRIP borehole (Fig. 9) where this signal strength was reached after more than 500 m. The reduction in observed moulin water pressure, indicative of 25m head of water draining out of the moulin, produced an increase in RSSI by around 10 dB. This confirms that the presence of liquid water increases the signal attenuation.
- The variation in signal is much greater when Cryoegg is in the "atmospheric pressure" region of the moulin rather than when it is in >1 m of water (Fig. 11). When Cryoegg is reporting pressure close to atmospheric pressure, it is being splashed by the water, or water is flowing smoothly past it. In this scenario the water flow will spin and agitate Cryoegg on the end of the rope, creating variation in signal level because of the antenna pattern. The turbulent flow of the water will also create everchanging levels of attenuation. However, once Cryoegg is below the water surface, the viscosity of the water will reduce its spinning and agitation, and the attenuation due to the water will be constant.

616 Glaciohydrological implications

617 The water pressure recorded by Cryoegg in the moulin steadily decreased during the 40 minutes that 618 it remained in the plunge pool (Fig. 11). Eventually, the water level dropped below the Cryoegg and it 619 returned to atmospheric pressure with the characteristic fluctuations in RSSI caused by splashing 620 water. We interpret this as dynamic drainage of the plunge pool over the afternoon, as water backed 621 up in the drainage system forces its way to the glacier bed. Similar pressure variations have been 622 previously observed in moulins (Iken, 1972; Röthlisberger, 1980; Holmlund and Hooke, 1983) and 623 demonstrate that the subglacial drainage system is not in equilibrium but constantly fluctuating 624 (Röthlisberger and Lang, 1987). The characteristic step-pool system develops if the moulin persists for 625 more than one season (Gulley, 2009); the artificial moulin was drilled directly to the end in 2018, but 626 by 2019 was 'kinked' and a plunge pool formed approximately 50 m above the bed.

627 Simultaneous salt tracing in a supraglacial stream feeding the moulin and within the moulin itself 628 shows (Fig. 12) that the moulin discharge was slightly higher than the stream discharge – unsurprising, 629 as the stream that we measured was not the sole supply of water feeding the moulin. This experiment 630 demonstrates Cryoegg's potential for measuring hydrological parameters in locations that are difficult 631 to access. Previously, moulin discharge has been estimated at the surface (either by field measurements or remote sensing), which masks the effect of water being stored within the vertical 632 633 column of the moulin itself (Werder and others, 2010). We show that it is possible to monitor 634 supraglacial discharge, the height of the stored water column within the moulin, and the moulin 635 discharge simultaneously and in real time, providing a valuable new approach for future studies of 636 glacier hydrology.

637 Comparison with other wireless subglacial probes

638 The most successful wireless subglacial probe for deep ice has been the WiSe system (Smeets and 639 others, 2012). This was demonstrated returning a signal through 2500 m of ice in Greenland. This 640 system operated at 30 MHz in order to benefit from lower free-space path loss, but at the expense of 641 making the antennas very large. The WiSe system suffered from some skywave interference affecting 642 signal reception, which is a particular issue at 30 MHz and below, and required use of a large (5 m 643 long) HB9CV type antenna to mitigate against it. WiSe required a 1 W (+30 dBm) transmitter to 644 communicate at depths of more than 2000 m, but it is not clear how much of this power was actually 645 radiated – the ferrite-loaded antenna used was likely to be very lossy.

The GlacsWeb system originally operated at 433 MHz (Martinez and others, 2004) but later (Martinez and others, 2013) used 151 MHz, giving a maximum reported range in ice of 70 m (Hart and others, 2019). Cryoegg is specifically designed for deep ice, and hence its radio performance greatly exceeds that of GlacsWeb, enabling its use in at least 1000 m of ice. Our probe and receiving antenna are both more compact than the WiSe system and we use commercially-available radio modules that adhere to an international open standard, which means that the key components are likely to be readily available well into the future.

653 CONCLUSION AND OUTLOOK

We have undertaken a full re-design of the wireless subglacial sensor platform Cryoegg, using a new radio link technology and improved link budget design, and demonstrate that it can transmit sensor data in real time through more than 1.3 km of cold ice. Deployments in moulins in temperate ice show that Cryoegg is a valuable tool for recording englacial and subglacial hydrological properties in situ, and hence giving further insight into processes in these environments. The EC sensor, originally intended as a proxy for total dissolved solids in subglacial water, can facilitate salt dilution gauging for

- real-time estimates of discharge. All sensors operated well, revealing englacial conditions and
 demonstrating their applicability for future subglacial deployments. The sensors fitted to the existing
 design were chosen because of their ease of implementation and their applicability to studies of
 subglacial hydrology, but Cryoegg can be adapted to support other sensors.
- 664 Future developments will refine and enhance the design, particularly with respect to the antenna 665 performance and mechanical design, so that we have a robust instrument capable of returning data 666 for months or years through 2.5 km of ice. This would enable us to match the performance of the WiSe system (Smeets and others, 2012) but with more compact antennas, enabling the Cryoegg to 'roam' 667 668 through englacial and subglacial hydrological systems to collect spatially and temporally distributed 669 measurements, reported in real time. Cryoegg technology will also be adapted for englacial studies in 670 irregular and refreezing hot-water-drilled boreholes, by creating a cylindrical form factor with a much 671 smaller diameter than Cryoegg, allowing deployment in a smaller borehole.

672 ACKNOWLEDGEMENTS

This work was funded by UK Engineering and Physical Sciences Research Council (EPSRC) NewInvestigator Award EP/R03530X/1, awarded to Liz Bagshaw.

- 675 Thanks to all who assisted us in the field, in particular: Trevor Popp, Romain Duphil and the EastGRIP 676 drill team; "Mickey" MacKay, Eliza Dawson and Rob Law at the RESPONDER site; Lai Bun Lok, Jonathan 677 Hawkins, the ETH Zürich VAW Glaciology team and the Swiss Air Force at the Rhône Glacier site. Phil 678 Anderson gave us use of the RF test chamber. We are grateful for in-kind contributions of equipment 679 from Quadro and Radiocrafts. Particular thanks to the team at DGF Engineering for the machining of 680 the caseworks, to Protronix EMS Ltd for the rapid production of the electronics, and to Justin Johnson 681 of InnovAntennas for the custom design of the receiving antenna. Miles Gould provided assistance 682 with the algebra for Equation 3. We would particularly like to thank our anonymous reviewers for their 683 helpful contributions and suggestions.
- 684 EastGRIP is directed and organized by the Center of Ice and Climate at the Niels Bohr Institute. It is 685 supported by funding agencies and institutions in Denmark (A. P. Møller Foundation, University of 686 Copenhagen), USA (US National Science Foundation, Office of Polar Programs), Germany (Alfred 687 Wegener Institute, Helmholtz Centre for Polar and Marine Research), Japan (National Institute of Polar 688 Research and Arctic Challenge for Sustainability), Norway (University of Bergen and Bergen Research 689 Foundation), Switzerland (Swiss National Science Foundation), France (French Polar Institute Paul-690 Emile Victor, Institute for Geosciences and Environmental research) and China (Chinese Academy of 691 Sciences and Beijing Normal University)
- Bryn Hubbard acknowledges support for hot-water borehole drilling from HEFCW through a CapitalEquipment Grant awarded to Aberystwyth University, and from the NERC through grant NE/K006126.
- The RESPONDER project is funded by a Consolidator Grant to Poul Christoffersen from the European
 Research Council, under the European Union's Horizon 2020 research and innovation programme
 (Grant 683043).
- Tom Chudley was supported by a Natural Environment Research Council Doctoral Training Partnership
 Studentship (Grant NE/L002507/1).
- ArcticDEM was created from DigitalGlobe, Inc., imagery and funded under National ScienceFoundation awards 1043681, 1559691, and 1542736
- 701

702 **REFERENCES**

703

Bagshaw E and 10 others (2018) Utilising Wireless Sensor Technologies in Glaciology. *Paper presented at AGU Fall Meeting 2018*. Washington, DC
 http://adsabs.harvard.edu/abs/2018AGUFM.C43D1823B

- Bagshaw EA and 6 others (2014) Novel wireless sensors for in situ measurement of sub-ice
 hydrologic systems. *Annals of Glaciology* 55(65), 41–50 (doi:10.3189/2014AoG65A007)
- Bagshaw EA and 7 others (2012) E-tracers: Development of a low cost wireless technique for
 exploring sub-surface hydrological systems. *Hydrological Processes* 26(20), 3157–3160
 (doi:10.1002/hyp.9451)
- Barrella T, Barwick S and Saltzberg D (2011) Ross Ice Shelf in situ radio-frequency ice attenuation.
 Journal of Glaciology 57(201), 61–66 (doi:10.3189/002214311795306691)
- Barwick S, Besson D, Gorham P and Saltzberg D (2005) South Polar in situ radio-frequency ice
 attenuation. *Journal of Glaciology* 51(173), 231–238 (doi:10.3189/172756505781829467)
- van den Broeke MR and 7 others (2016) On the recent contribution of the Greenland ice sheet to sea
 level change. *The Cryosphere* **10**(5), 1933–1946 (doi:10.5194/tc-10-1933-2016)
- Bullington K (1947) Radio Propagation at Frequencies above 30 Megacycles. *Proceedings of the Institute of Radio Engineers* 35(10), 1122–1136 (doi:10.1109/JRPROC.1947.232600)
- Chandler DM and 11 others (2013) Evolution of the subglacial drainage system beneath the
 Greenland Ice Sheet revealed by tracers. *Nature Geoscience* 6(3), 195–198
 (doi:10.1038/ngeo1737)
- Chu VW (2014) Greenland ice sheet hydrology: A review. *Progress in Physical Geography: Earth and Environment* 38(1), 19–54 (doi:10.1177/0309133313507075)
- Chudley TR and 6 others (2019) Supraglacial lake drainage at a fast-flowing Greenlandic outlet
 glacier. *Proceedings of the National Academy of Sciences* **116**(51), 25468–25477
 (doi:10.1073/pnas.1913685116)
- Church G, Bauder A, Grab M, Rabenstein L, Singh S and Maurer H (2019) Detecting and
 characterising an englacial conduit network within a temperate Swiss glacier using active
 seismic, ground penetrating radar and borehole analysis. *Annals of Glaciology* 60(79), 193–
 205 (doi:10.1017/aog.2019.19)
- Clarke GKC (1987) Subglacial till: A physical framework for its properties and processes. *Journal of Geophysical Research: Solid Earth* 92(B9), 9023–9036 (doi:10.1029/JB092iB09p09023)
- Dahl-Jensen D, Gundestrup N, Gogineni SP and Miller H (2003) Basal melt at NorthGRIP modeled
 from borehole, ice-core and radio-echo sounder observations. *Annals of Glaciology* 37, 207–
 212 (doi:10.3189/172756403781815492)
- Davies K (1990) *Ionospheric Radio (IEE Electromagnetic Waves Series, Vol. 31).* The Institution of
 Engineering and Technology, London, UK

- Doyle SH and 7 others (2018) Physical Conditions of Fast Glacier Flow: 1. Measurements From
 Boreholes Drilled to the Bed of Store Glacier, West Greenland. *Journal of Geophysical Research: Earth Surface* 123(2), 324–348 (doi:10.1002/2017JF004529)
- European Committee for Standardisation (2013) EN 13757-4:2013 Communication systems for
 meters and remote reading of meters. Wireless meter readout (Radio meter reading for
 operation in SRD bands). https://bsol.bsigroup.com/en/Bsol-Item-Detail Page/?pid=0000000030249978
- Flowers GE (2018) Hydrology and the future of the Greenland Ice Sheet. *Nature Communications*9(1), 2729 (doi:10.1038/s41467-018-05002-0)
- Friis HT (1946) A Note on a Simple Transmission Formula. *Proceedings of the Institute of Radio Engineers* 34(5), 254–256 (doi:10.1109/JRPROC.1946.234568)
- 750 Griffiths J (1987) Radio wave propagation and antennas. Prentice-Hall, Hemel Hempstead, UK
- Gulley J (2009) Structural control of englacial conduits in the temperate Matanuska Glacier, Alaska,
 USA. *Journal of Glaciology* 55(192), 681–690 (doi:10.3189/002214309789470860)
- Gundrestrup NS, Clausen HB and Hansen BL (1994) The UCPH borehole logger. *Memoirs of National Institute of Polar Research* Special Issue 49, 224–233 https://icedrill.org/library/ucph borehole-logger
- Hargreaves ND (1978) The Radio-Frequency Birefringence of Polar Ice. *Journal of Glaciology* 21(85),
 301–313 (doi:10.3189/S0022143000033499)
- Hart JK, Martinez K, Basford PJ, Clayton AI, Robson BA and Young DS (2019) Surface melt driven
 summer diurnal and winter multi-day stick-slip motion and till sedimentology. *Nature Communications* 10(1), 1–11 (doi:10.1038/s41467-019-09547-6)
- Hoffman MJ and 9 others (2016) Greenland subglacial drainage evolution regulated by weakly
 connected regions of the bed. *Nature Communications* 7(1), 1–12
 (doi:10.1038/ncomms13903)
- Hofstede C and 7 others (2018) Physical Conditions of Fast Glacier Flow: 2. Variable Extent of
 Anisotropic Ice and Soft Basal Sediment From Seismic Reflection Data Acquired on Store
 Glacier, West Greenland. *Journal of Geophysical Research: Earth Surface* 123(2), 349–362
 (doi:10.1002/2017JF004297)
- Holmlund P and Hooke RLeb (1983) High Water-Pressure Events in Moulins, Storglaciären, Sweden. *Geografiska Annaler: Series A, Physical Geography* 65(1–2), 19–25
 (doi:10.1080/04353676.1983.11880070)
- Hubbard BP, Sharp MJ, Willis IC, Nielsen MK and Smart CC (1995) Borehole water-level variations
 and the structure of the subglacial hydrological system of Haut Glacier d'Arolla, Valais,
 Switzerland. *Journal of Glaciology* 41(139), 572–583 (doi:10.3189/S0022143000034894)
- Iken A (1981) The Effect of the Subglacial Water Pressure on the Sliding Velocity of a Glacier in an
 Idealized Numerical Model. *Journal of Glaciology* 27(97), 407–421
 (doi:10.3189/S0022143000011448)

- Iken A (1972) Measurements of Water Pressure in Moulins as Part of a Movement Study of the
 White Glacier, Axel Heiberg Island, Northwest Territories, Canada. *Journal of Glaciology* 11(61), 53–58 (doi:10.3189/S0022143000022486)
- Iken A and Bindschadler RA (1986) Combined measurements of Subglacial Water Pressure and
 Surface Velocity of Findelengletscher, Switzerland: Conclusions about Drainage System and
 Sliding Mechanism. *Journal of Glaciology* **32**(110), 101–119
 (doi:10.3189/S002214300006936)
- Iken A, Echelmeyer K, Harrison W and Funk M (1993) Mechanisms of fast flow in Jakobshavns Isbræ,
 West Greenland: Part I. Measurements of temperature and water level in deep boreholes.
 Journal of Glaciology **39**(131), 15–25 (doi:10.3189/S0022143000015689)
- Joughin I (2017) MEaSUREs Greenland Ice Velocity Annual Mosaics from SAR and Landsat, Version 1.
 (doi:10.5067/OBXCG75U7540)
- Joughin I, Smith BE, Howat IM, Scambos T and Moon T (2010) Greenland flow variability from ice sheet-wide velocity mapping. *Journal of Glaciology* 56(197), 415–430
 (doi:10.3189/002214310792447734)
- Kamb B (1987) Glacier surge mechanism based on linked cavity configuration of the basal water
 conduit system. *Journal of Geophysical Research: Solid Earth* 92(B9), 9083–9100
 (doi:10.1029/JB092iB09p09083)
- Kamb B (1970) Sliding motion of glaciers: Theory and observation. *Reviews of Geophysics* 8(4), 673–
 728 (doi:10.1029/RG008i004p00673)
- Karlsson NB and Dahl-Jensen D (2015) Response of the large-scale subglacial drainage system of
 Northeast Greenland to surface elevation changes. *The Cryosphere* 9(4), 1465–1479
 (doi:10.5194/tc-9-1465-2015)
- Lewis C and 6 others (2015) Airborne fine-resolution UHF radar: an approach to the study of
 englacial reflections, firn compaction and ice attenuation rates. *Journal of Glaciology* 61(225), 89–100 (doi:10.3189/2015JoG14J089)
- Li J and 8 others (2018) Multi-channel and multi-polarization radar measurements around the NEEM
 site. *The Cryosphere* 12(8), 2689–2705 (doi:10.5194/tc-12-2689-2018)
- MacGregor JA and 11 others (2015) Radar attenuation and temperature within the Greenland Ice
 Sheet. Journal of Geophysical Research: Earth Surface 120(6), 983–1008
 (doi:10.1002/2014JF003418)
- MacGregor JA, Winebrenner DP, Conway H, Matsuoka K, Mayewski PA and Clow GD (2007)
 Modeling englacial radar attenuation at Siple Dome, West Antarctica, using ice chemistry
 and temperature data. *Journal of Geophysical Research: Earth Surface* 112(F3)
 (doi:10.1029/2006JF000717)
- Martinez K, Basford PJ, Jager DD and Hart JK (2013) Using a heterogeneous sensor network to
 monitor glacial movement. *Paper presented at 10th European Conference on Wireless Sensor Networks*. http://eprints.soton.ac.uk/348427/1/paper.pdf

- Martinez K, Ong R and Hart J (2004) Glacsweb: a sensor network for hostile environments. 2004 First
 Annual IEEE Communications Society Conference on Sensor and Ad Hoc Communications and
 Networks, 2004. IEEE SECON 2004. 81–87 (doi:10.1109/SAHCN.2004.1381905)
- 818 Moore RD (2005) Introduction to salt dilution gauging for streamflow measurement Part 3: Slug 819 injection using salt in solution. *Streamline Watershed Management Bulletin* **8**(2), 1–6
- Morlighem M and 31 others (2017) BedMachine v3: Complete Bed Topography and Ocean
 Bathymetry Mapping of Greenland From Multibeam Echo Sounding Combined With Mass
 Conservation. *Geophysical Research Letters* 44(21), 11,051-11,061
 (doi:10.1002/2017GL074954)
- Ng FSL (2000) Canals under sediment-based ice sheets. *Annals of Glaciology* **30**, 146–152
 (doi:10.3189/172756400781820633)
- Nienow PW, Sole AJ, Slater DA and Cowton TR (2017) Recent Advances in Our Understanding of the
 Role of Meltwater in the Greenland Ice Sheet System. *Current Climate Change Reports* 3(4),
 330–344 (doi:10.1007/s40641-017-0083-9)
- Nienow PW and 6 others (2005) Hydrological controls on diurnal ice flow variability in valley glaciers.
 Journal of Geophysical Research: Earth Surface 110(F4) (doi:10.1029/2003JF000112)
- Nye JF (1976) Water Flow in Glaciers: Jökulhlaups, Tunnels and Veins. *Journal of Glaciology* 17(76),
 181–207 (doi:10.3189/S002214300001354X)
- Plewes LA and Hubbard B (2001) A review of the use of radio-echo sounding in glaciology: *Progress in Physical Geography* (doi:10.1177/030913330102500203)
- Pritchard HD, Arthern RJ, Vaughan DG and Edwards LA (2009) Extensive dynamic thinning on the
 margins of the Greenland and Antarctic ice sheets. *Nature* 461(7266), 971–975
 (doi:10.1038/nature08471)
- 838 RC1701xx-MBUS Datasheet (2018) Radiocrafts AS, Oslo, Norway
 839 https://radiocrafts.com/uploads/rc1701xx-mbus_datasheet.pdf
- Röthlisberger H (1980) Gletscherbewegung und Wasserabfluss. *Wasser, Energie, Luft* 72(9), 290–294
- Röthlisberger H (1972) Water Pressure in Intra- and Subglacial Channels. *Journal of Glaciology* 11(62), 177–203 (doi:10.3189/S0022143000022188)
- Röthlisberger H and Lang H (1987) Glacial Hydrology. *Glacio-Fluvial Sediment Transfer: An Alpine Perspective*. Wiley, 207–84
- Schoof C (2010) Ice-sheet acceleration driven by melt supply variability. *Nature* 468(7325), 803–806
 (doi:10.1038/nature09618)
- Sheldon SG, Steffensen JP, Hansen SB, Popp TJ and Johnsen SJ (2014) The investigation and
 experience of using ESTISOL[™] 240 and COASOL[™] for ice-core drilling. *Annals of Glaciology* 55(68), 219–232 (doi:10.3189/2014AoG68A036)
- Smeets CJPP and 6 others (2012) A wireless subglacial probe for deep ice applications. *Journal of Glaciology* 58(211), 841–848 (doi:10.3189/2012JoG11J130)

- Sole AJ and 6 others (2011) Seasonal speedup of a Greenland marine-terminating outlet glacier
 forced by surface melt-induced changes in subglacial hydrology. *Journal of Geophysical Research: Earth Surface* **116**(F3) (doi:10.1029/2010JF001948)
- Stott J (2005) BBC R&D White Paper WHP 109: Digital Radio Mondiale: DRM, digital radio on long,
 medium and short waves another radio revolution?
 http://downloads.bbc.co.uk/rd/pubs/wbp./wbp.pdf files/WHP100.pdf
- 857 http://downloads.bbc.co.uk/rd/pubs/whp/whp-pdf-files/WHP109.pdf
- Tedstone AJ, Nienow PW, Gourmelen N, Dehecq A, Goldberg D and Hanna E (2015) Decadal
 slowdown of a land-terminating sector of the Greenland Ice Sheet despite warming. *Nature*526(7575), 692–695 (doi:10.1038/nature15722)
- Tranter M, Sharp MJ, Lamb HR, Brown GH, Hubbard BP and Willis IC (2002) Geochemical weathering
 at the bed of Haut Glacier d'Arolla, Switzerland—a new model. *Hydrological Processes* 16(5),
 959–993 (doi:10.1002/hyp.309)
- UM10204 I2C-bus specification and user manual (2014) NXP Semiconductors, Eindhoven,
 Netherlands https://www.nxp.com/docs/en/user-guide/UM10204.pdf
- Walder JS (1986) Hydraulics of Subglacial Cavities. *Journal of Glaciology* 32(112), 439–445
 (doi:10.3189/S0022143000012156)
- Werder MA, Schuler TV and Funk M (2010) Short term variations of tracer transit speed on alpine
 glaciers. *The Cryosphere* 4(3), 381–396 (doi:10.5194/tc-4-381-2010)
- Young TJ and 8 others (2018) Resolving the internal and basal geometry of ice masses using imaging
 phase-sensitive radar. *Journal of Glaciology* 64(246), 649–660 (doi:10.1017/jog.2018.54)
- 872

874 APPENDIX A – BATTERY LIFE CALCULATION

875 Table A1 – battery life calculation

876

Activity	Quantity	Unit
Energy consumption during measure and transmit	0.5	J
Duration of measure and transmit	3.2	S
Current consumption during sleep	500	nA
Nominal battery voltage	3.7	V
Power consumption during sleep	1.85	uW
Battery nominal voltage	3.7	V
Battery nominal capacity	400	mAh
Battery nominal capacity	1.48	Wh
Battery nominal capacity	5328	J
Derate factor for operating in the cold	50%	
Battery effective capacity in the field	2664	J
For 2 measurements per day:		
Active measurement time per day:	6.4	seconds
Sleep time per day	86393.6	seconds
Energy used whilst active	1	J
Energy used whilst asleep	0.16	J
Total energy used per day	1.16	J
Number of days the system will run for	2296	days
Number of years the system will run for	6.3	years